

Blook™

LED Transilluminator



BLook™ LED transilluminator

Catalog number: BK001

Description

BLook™ is an extraordinary blue light LED transilluminator for the detection of nucleic acids or protein under non-UV conditions. The wavelength of the special blue LED lights is 470 nm (fig 1), hence no damage to your nucleic acids or protein. Since UV is not used, there is no need for any special personal eye or skin protection. The blue LED lights are arranged under the viewing area (200 × 120 mm). An amber filter, on hinges, is lowered into position once your gel is mounted. The stained gel is now ready for observation. This instrument has a specially designed ergonomic 4° angle, so users can easily sit on a chair to see the experiment results. BLook™ is designed to inspect the gel after electrophoresis and stained with the Novel Juice, Novel Green, Novel Green *Plus*. Further, it is perfectly designed for OnePCR™, OnePCR™ HiFi, OnePCR™ HotStar, OnePCR™ *Plus*, OneMARK B, and OneMARK 100, which contains the fluorescent stain compatible with the blue light wavelength. However, BLook™ is not suitable for ethidium bromide staining.




Fig.1. Emission spectrum for the BLook™ LED transilluminator

Features

- Ergonomic fusion-Patented 4° ergonomic viewing angle (“Golden Angle”).
- Optimized for use with the nucleic acid and protein fluorescent dyes.
- Blue light source good for 30,000 hours.
- No risk of UV damage for high quality work experience.
- Smart power-saving function - Automatic power shut-off option at 5 minutes.
- Gel-cutting knife - Cut out the target from the gel for further experiment.
- A mini darkroom for viewing gel or placing the photographic equipment.

Specifications


Unit dimensions (W × L × H)	295 × 215 × 42 mm (11.6 × 8.4 × 1.6 inch)
Gel viewing dimensions (W × L × H)	200 × 120 mm (7.8 × 4.7 inch)
Weight (g)	1280
Input voltage	100-240Vac, 50 / 60 Hz
Input current	1.4A
LED source	Built-in blue light LED module
LED life (hours)	>30,000
Emission maxima (nm)	470
Store temperature	25°C
Operating temperature	Ambient to 40°C
Auto shut-off (minute)	5
Filter type :	Amber filter (ideal for Novel Juice, Novel Green, Novel Green <i>Plus</i> , OnePCR™, OnePCR™ HiFi, OnePCR™ HotStar, OnePCR™ <i>Plus</i> , OneMark B, OneMARK 100)

Contents:



- (A) Power Cord: Used to connect the BLook™ port with the power socket.
 (B) BLook™ LED transilluminator
 (C) Gel-Cutting Knife: Cut out the target from the gel for further experiment.
 (D) Replacement Blade: Used as the new blade.

BLook™ LED transilluminator



Operating

1. Place the BLook™ LED transilluminator on a level bench, with enough space around the device for allowing air circulation and preventing overheating.



2. Connect the power cord and the socket.



3. Open the amber lid.

Note: Please open amber lid carefully by two hands.

4. Place the gel on glass area.

Note: The automatic power shut-off option at 5 minutes. Once the LED lights are off, wait for 3 minutes, and then press the power button again to re-turn on the light.

5. Close the amber lid above the gel.

6. Press ON/OFF switch to ON mode.



7. Now the fluorescent samples can be viewed or cut.

Note: For cutting fluorescent samples out of the gel, please lift the lid to access the gel.



8. When finished, turn the transilluminator off.

9. Clean the transilluminator filter and surface by wiping with a wet, not scratching cloth.

Troubleshooting

Refer to the table below to troubleshoot problems that you may encounter when using with the BLook™ LED transilluminator.

Problem	Cause	Solution
Low sensitivity	Wavelength may not be right	Check the excitation and emission wavelengths of the fluorescence dye
Automatic shut-off	Automatic power-off system	Please wait for 3 minutes, then press the power button again to re-turn on the light
High background, suboptimal, or no image	Low sample concentration	Check the concentration of sample

Related Ordering Information

Cat. No.	Description	Size
LD001-1000	Novel Juice	1 ml
LD002-0500	Novel Green	500 µl
LD003-0500	Novel Green <i>Plus</i>	500 µl
MB203-0100	OnePCR™	100 Reactions
MB205-0100	OnePCR™ HiFi	100 Reactions
MB206-0100	OnePCR™ HotStar	100 Reactions
MB207-0100	OnePCR™ <i>Plus</i>	100 Reactions
DM101-0100	OneMARK 100	600 µl
DM110-0100	OneMARK B	600 µl
SM213-0250	amaR OnePCR	250 rxn

Note: Please, read the installation instruction carefully before installing the BLook™ LED transilluminator.

This instrument is intended for clinical and research laboratory use and it must be operated only by specialized personnel aware of the potential risks associated with the chemical and biological agents normally used with this unit. This instrument is meant for use only by specialized personnel that know the health risks associated with blue light radiation and with reagents that are normally used with this instrument. It is designed to shield only the person working in front of the system and to observe the gel fluorescence emission clearly.

Maintenance and Repair

Cleaning and Maintenance

The transilluminator should be damply wiped off with a soft cotton cloth. The cloth should not be too wet. Do not use aggressive cleaning detergents or solvents. After use of the transilluminator, we recommend to wipe off the frame and the filter glass with a damp cloth. Take care that the cloth is not soaked with scrubbing or corrosive detergents as this could damage the filter glass. We suggest using alcohol, glass cleaners or similar mild detergents.

Pay attention to wearing appropriate gloves when cleaning areas (as the filter plate of the transilluminator, switches, protection shield) which may have been in contact with carcinogenic or toxic reagents. The protection shield is made of tempered glass and must only be cleaned with a damp cloth.

Method of Replacing with the New Blade

When replacing with the new blade, unscrew the black handling part of the knife counterclockwise from the blue part. Push out the blue part's blade-holding section. Remove the old blade, and open the "SPARE BLADE" for taking out the new replacement blade. Insert it into the X-shaped blade insertion hole. Restore the blue connector back into the original position, and then screw tight the blue part with the black handling part clockwise.

Caution: The blade is quite sharp. Please act with caution to avoid cuts.

Replacement of Spare Parts

Only original spare parts are allowed.

Servicing

Please contact your local dealer.

The BLook™ LED transilluminator is warranted against defects in materials and workmanship for 1 year. If any defects occur in the instrument or accessories during this warrant period, Simply Biologics, Inc. will repair or replace the defective parts at its discretion without charge. The following defects, however, are specifically excluded:

1. Defects caused by improper operation.
2. Repair or modification done by anyone other than Simply Biologics, Inc. or an authorized agent.
3. Damage caused by substituting alternative parts.
4. Use of fittings or spare parts supplied by anyone other than Simply Biologics, Inc.
5. Damage caused by accident or misuse.
6. Damage caused by disaster.
7. Corrosion caused by improper solvent or sample.

For any inquiry or request for repair service, contact your local Simply Biologics, Inc. office. Inform the model and serial number of your instrument.

REGULATORY NOTICE

IMPORTANT: This Simply Biologics, Inc. instrument is designed and certified to meet safety standards. Certified products are safe to use when operated in accordance with the instruction manual. This instrument should not be modified or altered in any way. Alteration of this instrument will:

1. Void the manufacturer's warranty
2. Void the safety and EMC certification
3. Create a potential safety hazard

Simply Biologics, Inc. is not responsible for any injury or damage caused by the use of this instrument for purposes other than those for which it is intended, or by modifications of the instrument not performed by Simply Biologics, Inc. or an authorized agent.

Appendix

Explanation of Symbols and Warnings



BLook™ LED transilluminator complies with the European Community Safety requirements.

A CE Marking is a European marking of conformity that indicates a product complies with the essential requirements of the applicable European laws or directives with respect to safety, health, and environment and consumer protection. Generally, this conformity to the applicable directives is done through self-declaration. The CE Marking is required on products in the countries of the European Economic Area (EEA) to facilitate trade among the member countries. The manufacturer or their authorized representative established in the EEA is responsible for affixing the CE Marking to their product. The CE Marking provides a means for a manufacturer to demonstrate that a product complies with a common set of laws required by all of the countries in the EEA to allow free movement of trade within the EEA countries.



Caution

Caution risk of danger.

Consult the User Guide for further safety information.

Do not dispose of this product in unsorted municipal waste.



Caution

To minimize negative environmental impact from disposal of electronic waste, do not dispose of electronic waste in unsorted municipal waste.

Follow local municipal waste ordinances for proper disposal provision and contact customer service for information about responsible disposal options.

**CAUTION**

LED radiation.
DO NOT STARE INTO
BEAM.
Class 2 LED product.

Caution

The BLook™ LED transilluminator is classified as a Class 2 LED product, which is indicated by the symbol to the left, radiation when open, do not stare into the beam

Safety Information

The BLook™ LED transilluminator is an electrical device. Never touch the power cord or outlet with wet hands. Do not use this device in damp areas or while standing on damp floors. Do not attempt to open the BLook™ LED transilluminator. The BLook™ LED transilluminator should be used with the supplied power cord. Only these power cords should be used to power the device. Attach the power cord to the BLook™ LED transilluminator at the back of the device. Plug the other end of the power cord into a properly grounded electrical outlet, ensuring the correct plug adaptor is attached. Always disconnect the BLook™ LED transilluminator from the electrical outlet before cleaning the device.

The BLook™ LED transilluminator does not produce UV- light; however, it does utilize an intense blue light for viewing gels. It should be noted that published literature has identified blue light as a possible risk factor for macular degeneration; however, no clinical studies have been published. Therefore, the BLook™ LED transilluminator amber filter unit provided with this device should be used to protect your eyes while viewing gels. The amber filter unit is NOT a safety screen for UV emission, and will NOT protect your eyes when viewing gels on UV transilluminators. After viewing and documenting the gel or sample, always press the unit off.

1.Manufacturer:

Taqkey Science Co., Ltd.

1F., No. 60, Jiabei 2nd St., Jhunan Township, Miaoli County, TAIWAN

Tel: +886-37-625 816, +886-2-2577 7306

E-mail: service@taqkey.com

- 2.Please position and use the LED transilluminator on the level laboratory benchtop.
- 3.If the LED transilluminator is positioned on the unlevelled laboratory benchtop or platform, it may cause the sample gel to slide, thus making the user unable to observe the gel.
- 4.After the LED transilluminator is switched on for 5 minutes, the automatic disconnecting device will automatically switch off the power source. Please wait for 3 minutes before switching on the device again. If the user doesn't wait for 3 minutes and continuously presses the ON/OFF button, it may cause the device to be overheated and dysfunctional as a result.

Warranty

Simply Biologics, Inc. warrants that BLook™ LED transilluminator will be free from defects in material and workmanship for a period of one year from date of purchase. If a defect is present, Simply Biologics, Inc. will, at its option, repair, replace, or refund the purchase price of this product at no charge to you, provided it is returned during the warranty period. This warranty does not apply if the product has been damaged by accident, abuse, misuse or misapplication, or from ordinary wear and tear. This warranty shall be limited to the replacement of defective products. It is expressly agreed that this warranty will be in lieu of all warranties of fitness and in lieu of the warranty of merchantability.

